MERCATOR MEDICAL



User instruction

The instruction below should be used in conjunction with detailed information on the packaging.

Short description of the product

Nitrile examination and protective gloves, powder-free, non-sterile for disposable use

Full description of the product

Raw material : nitrile

External surface : bisque with fingertip textured, polymerized

Internal surface : polymerized + chlorinated

Cuff : beaded

Colour : blue/white/violet blue/black/cobalt

Shape : ambidextrous, fitting to the right and left hand Size range : XS (5-6), S (6-7), M (7-8), L (8-9), XL (9-10)

AQL : 1.5

Quantity in packaging : 50/100/200 pcs. by weight

Shelf life : 3 years (from the date of manufacturing)

Storage instructions

It is recommended to store the gloves in dry place, in the temperature of 5-35°C and to protect them against direct sunlight and fluorescent light. Recommended relative humidity in the room where the gloves are stored is 60 $\pm 20\%$.

Keep the gloves in a distance of not less than 1m from heating devices, sources of fire and ozone.

Do not keep in direct vicinity of solvents, oils, fuels and lubricants.

Food contact

Gloves are marked with food contact symbol and comply with the requirements of Regulation (EU) No 10/2011, European Regulation (EC) No 1935/2004 and with Regulation (EC) No 2023/2006 on Good Manufacturing Practice. Gloves are suitable for handling any type of food and have been tested for Overall Migration Test acc. EN 1186:

Extraction conditions	Analysis results	Test Result
(tested for 2 h in 40°C)	[mg/dm ²]	(limit < 10 mg/dm²)
3% acetic acid	0.8	Pass
10% ethanol	ND	Pass
Olive oil	ND	Pass

MDR classification & compliance

Gloves are classified as class I Medical Device as per Annex VIII of the REGULATION (EU) 2017/745 and comply to standards: EN 455-1:2000, EN 455-2:2015, EN 455-3:2015, EN 455-4:2009, EN ISO 15223-1:2016, EN 1041:2008+A1:2013.

Authorized Representative

Mercator Medical S.A. Address: ul. H. Modrzejewskiej 30 31-327 Krakow, Poland

PPE classification & compliance

Gloves are category III Personal Protective Equipment as per Annex I of the Regulation 2016/425 and comply to standards: EN 420:2003+A1:2009, EN ISO 374-1:2016 (Type B), EN 374-2:2014, EN 16523-1:2015, EN 374-4:2013, EN ISO 374-5:2016.

EU Type Examination Certificate issued by: SATRA (Notified Body No. 2777) Checking of PPE manufactured:

SATRA Technology Europe Limited Bracetown Business Park Clonee. D15YN2P Republic of Ireland

Intended use

These are non-sterile examination and protective gloves for single use, intended for use in medical field to: protect patient and user from cross-contamination, conducting medical examinations, diagnostic and therapeutic procedures and for handling medical contaminated material. Gloves are classified as Medical Devices Class I and as a Personal Protective Equipment category III. Their design and labelling corresponds to the requirements of the European Medical Device Regulation (EU) 2017/745 and the European Regulation 2016/425 on Personal Protective Equipment. Gloves should be used solely according to their intended application.

Precautions and indications for use

Dry hands before putting the gloves on. Before usage, inspect the gloves for any defect or imperfections. Use at least 1 pair of gloves for one patient and one procedure, these are disposable gloves. Do not let chemical substances get under the gloves through the cuff. If a chemical substance reaches the skin, wash it away immediately with plenty of water with soap. If the gloves get punctured, torn or broken during their use, take them off and put on the new ones. Avoid using gloves dirty in the inside as they may cause irritation leading to skin inflammation or more serious damages. The gloves should not be used in contact with open fire and to protect against any sharp tools. The gloves are not intended for welding, electric shock protection, ionizing radiation or from the effect of hot or cold objects.

This information does not reflect the actual duration of protection in the workplace and the differentiation between mixtures and pure chemicals.

The chemical penetration resistance has been assessed under laboratory conditions from samples taken from the palm only (except in case where glove is equal to or over 400 mm – where the cuff is tested also) and relates only to the chemical tested and to the tested specimen. It can be different if the chemical is used in a mixture.

It is recommended to check that the gloves are suitable for the intended use because the conditions at the workplace may differ from the type test depending on the temperature, abrasion and degradation.

When used, protective gloves may provide less resistance to the dangerous chemical due to changes in physical properties. Movements, snagging, rubbing, degradation caused by the chemical contact etc. may reduce the actual use time significantly. For corrosive chemicals, degradation can be the most important factor to consider in selection of chemical resistant gloves.

Gloves are suitable for special purposes as they are examination gloves where risk of injury to the wrist is considered to be minimal, gloves are shorter than EN 420 min. length requirement.

Components / hazardous components

Some gloves may contain components known to be a possible cause of allergy for person allergic to them, who may develop contact irritation and/or allergic reaction. In case of an allergic reaction, seek medical assistance immediately.

Disposal

Used gloves can be contaminated with contagious or other hazardous substances. They should be disposed of in accordance with local regulation. Gloves should be buried or burned under controlled conditions.

Manufacturer

MERCATOR MEDICAL (Thailand)Ltd. 88/8 Moo 12 Tambon Kampaengphet, Amphur Rattaphum,Songkhla 90180,Thailand.

www.mercator.co.th

MERCATOR MEDICAL



Permeation performance levels as per Level1>10min , Level2>30 min, level3>60 min, level4>12	EN 374-4:2013	
Test results acc. to EN 1652		
Chemical	Level	Degradation [%]
35% Ethanol	6	55.0
40% Isopropanol	6	68.7
10 Acetic acid	4	53.5
50 % Benzalkonium chloride*	6	29.5
4% Chlorhexidine digluconate **	6	32.9
10% Phosphoric acid	6	14.0
40 % Sodium hydroxide	6	2.6
12 % Sodium hypochlorite	6	22.7
50% Sulphuric Acid	6	21.1
5% Ethidium Bromide	6	32.9
3% Hydrogen peroxide	6	44.0
30% Hydrogen peroxide	2	52.8
37% Formaldehyde	5	20.0
50% Glutaraldehyde	6	22.9
0.1% Phenol	6	24.7

- Permeation rate 5 g/cm2 /min,EN374-4:2013 degradation levels indicate the change in puncture resistance of the gloves after exposure to the challenge chemical.
- ** Permeation rate 7 g/cm2 /min,EN374-4:2013 degradation levels indicate the change in puncture resistance of the gloves after exposure to the challenge chemical.

Noted: 1) Glove minimum length for Lab application accordance to EN455-2

Test acc. To EN 374-2:2014 – Level 2 (ISO 2859)		Test acc. To EN ISO 374-5:2016				
	Performance level	AQL	Protection against bacteria & fungi	Pass		
	Level 3	< 0.65	Protection against viruses	Pass		
	Level 2	<1.5				
	Level 1	< 4.0				


Symbols used on the packaging Do not re-use / gloves are Non-sterile gloves Powdered gloves intended for single use Do not use, if package is Keep away from solar and Powder free gloves damaged fluorescent light Temperature limitation / Keep away from moisture, Presence of polymer coating on gloves store in temperature 5store in a dry place the inner surface of the glove 35°C Presence of cosmetic coating on Raw material – natural rubber Keep away from ozone latex the inner surface of glove Gloves with incorporated singlet Catalogue number Lot / batch number oxygen layer. EU Authorized Representative, symbol should be accompanied Presence of external texture on Expiry date by name and address of the glove Authorized Representative ISO 374-5:2016 Gloves protecting against Marking of gloves protecting NITRYL chemical dangers with digit Gloves made from nitrile against bacteria and fungi. literal odes Marking of gloves protecting against viruses, bacteria and Antistatic gloves Gloves made from vinyl fungi. Marking o type A chemical resistant gloves. Six tested chemicals shall be Date of manufacture Gloves made from neoprene identified by their code letter under pictogram. Marking o type B chemical resistant Manufacturer, symbol should gloves. Three tested chemicals shall be accompanied by name and Gloves made from polyisoprene be identified by their code letter address of Manufacturer under pictogram.


MERCATOR MEDICAL











risk (if applicable accompanied by 4 digit code of relevant performance

Food contact symbol (article is suitable for food contact, for details check the instruction for use)

Indicates compliance with the requirements of Russian market



Consult instructions for use

Package made from paper,



100

by weight

50 gloves by weight





Package is treated as municipal waste

qualify for recycling

200

200 gloves by weight

100 gloves by weight



Indicates compliance with the requirements of Ukrainian market



Additional information on inner side of package

Mr. Praneet Inthajak

Snr. QA/RA Manager

MERCATOR MEDICAL (Thailand) Ltd.



■ HOW TO PUT THE GLOVES ON? ■













■ HOW TO TAKE THE GLOVES OFF? ■











